

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005739.D
 Acq On : 01 Nov 2018 19:38
 Operator : JC/MD
 Sample : PB114370ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114370ZHE#05

Quant Time: Nov 02 04:33:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	118748	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	191707	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	170517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	71940	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	95137	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
35) Dibromofluoromethane	5.50	113	67847	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
50) Toluene-d8	8.71	98	239475	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.14	95	75134	43.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.46%	
Target Compounds						
16) Acetone	2.45	43	10946	3.057	ug/l	97
18) Methyl Acetate	2.78	43	5603	2.249	ug/l	96
20) Methylene Chloride	2.85	84	4957	3.311	ug/l	92
43) Isopropyl Acetate	6.46	43	103202	27.287	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005739.D
Acq On : 01 Nov 2018 19:38
Operator : JC/MD
Sample : PB114370ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114370ZHE#05

Quant Time: Nov 02 04:33:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

