

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005740.D
 Acq On : 01 Nov 2018 20:05
 Operator : JC/MD
 Sample : PB114370ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114370ZHE#06

Quant Time: Nov 02 04:35:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	123866	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	200244	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	174130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	73352	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	95114	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
35) Dibromofluoromethane	5.51	113	66939	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
50) Toluene-d8	8.71	98	252889	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	11.14	95	77554	43.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.42%	
Target Compounds						
16) Acetone	2.45	43	11736	3.412	ug/l	96
18) Methyl Acetate	2.78	43	6893	2.652	ug/l #	89
20) Methylene Chloride	2.85	84	4698	3.008	ug/l	91
43) Isopropyl Acetate	6.46	43	100021	25.319	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005740.D
Acq On : 01 Nov 2018 20:05
Operator : JC/MD
Sample : PB114370ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114370ZHE#06

Quant Time: Nov 02 04:35:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

