

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005741.D
 Acq On : 01 Nov 2018 20:32
 Operator : JC/MD
 Sample : PB114370ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114370ZHE#07

Quant Time: Nov 02 04:37:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	120669	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	197511	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	172813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	73223	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	89120	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
35) Dibromofluoromethane	5.50	113	70270	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
50) Toluene-d8	8.71	98	244806	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
62) 4-Bromofluorobenzene	11.14	95	76613	43.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.56%	
Target Compounds						
16) Acetone	2.45	43	11496	3.484	ug/l	96
18) Methyl Acetate	2.78	43	6571	2.595	ug/l	95
20) Methylene Chloride	2.86	84	2677	1.760	ug/l	89
25) 2-Butanone	4.71	43	2124	1.427	ug/l	100
43) Isopropyl Acetate	6.46	43	99967	25.655	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005741.D
Acq On : 01 Nov 2018 20:32
Operator : JC/MD
Sample : PB114370ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114370ZHE#07

Quant Time: Nov 02 04:37:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

