

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110119\
 Data File : VX013303.D
 Acq On : 01 Nov 2019 12:00
 Operator : JC/SP
 Sample : K5626-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S1TB2-191030

Quant Time: Nov 04 00:51:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	741626	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1118941	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	982479	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	450661	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	425064	48.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.76%	
35) Dibromofluoromethane	5.49	113	331339	47.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.42%	
50) Toluene-d8	8.71	98	1311804	49.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.13	95	429343	44.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110119\
Data File : VX013303.D
Acq On : 01 Nov 2019 12:00
Operator : JC/SP
Sample : K5626-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S1TB2-191030

Quant Time: Nov 04 00:51:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

