

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110119\
 Data File : VX013306.D
 Acq On : 01 Nov 2019 13:10
 Operator : JC/SP
 Sample : K5655-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 04 01:05:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	735641	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1109840	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	977734	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	447068	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	412674	47.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.70%	
35) Dibromofluoromethane	5.49	113	328810	47.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.48%	
50) Toluene-d8	8.71	98	1312417	50.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.02%	
62) 4-Bromofluorobenzene	11.13	95	434316	45.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110119\
Data File : VX013306.D
Acq On : 01 Nov 2019 13:10
Operator : JC/SP
Sample : K5655-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 04 01:05:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

