

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110119\
 Data File : VX013315.D
 Acq On : 01 Nov 2019 16:43
 Operator : JC/SP
 Sample : K5651-02DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS003MW326R-191031DL

Quant Time: Nov 04 02:14:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	754368	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1148182	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1022976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	483532	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	432110	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
35) Dibromofluoromethane	5.48	113	341167	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
50) Toluene-d8	8.71	98	1372545	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	459490	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
Target Compounds						
11) Tert butyl alcohol	3.00	59	17652	7.494	ug/l #	85
27) cis-1,2-Dichloroethene	4.58	96	14982	1.775	ug/l	81
40) Benzene	6.13	78	37620	1.226	ug/l	97
65) Chlorobenzene	10.14	112	1027699	49.622	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	84474	5.545	ug/l	99
88) 1,4-Dichlorobenzene	12.09	146	284932	18.357	ug/l	99
91) 1,2-Dichlorobenzene	12.38	146	2341298	153.748	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110119\
Data File : VX013315.D
Acq On : 01 Nov 2019 16:43
Operator : JC/SP
Sample : K5651-02DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS003MW326R-191031DL

Quant Time: Nov 04 02:14:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

