

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110121\
 Data File : VX025030.D
 Acq On : 01 Nov 2021 11:27
 Operator : JC/MD
 Sample : VX1101MBL06
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 02 03:30:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 02 03:29:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	179812	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	10.055	117	162702	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	74519	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	34728	30.298	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	60.600%
7) Chloroethane-d5	1.672	69	27003	34.466	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	68.940%#
11) 1,1-Dichloroethene-d2	2.306	63	70840	25.081	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	50.160%#
21) 2-Butanone-d5	4.458	46	97421	85.093	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.090%
24) Chloroform-d	5.062	84	122126	40.720	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.440%
26) 1,2-Dichloroethane-d4	5.952	65	87209	39.652	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.300%
32) Benzene-d6	5.976	84	185479	40.164	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.320%
36) 1,2-Dichloropropane-d6	7.312	67	66947	42.538	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.080%
41) Toluene-d8	8.653	98	172871	39.128	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.260%#
43) trans-1,3-Dichloroprop...	8.952	79	38047	42.067	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.140%
47) 2-Hexanone-d5	9.384	63	72356	90.300	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.300%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	97853	45.980	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.960%
66) 1,2-Dichlorobenzene-d4	12.323	152	66636	46.083	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110121\
Data File : VX025030.D
Acq On : 01 Nov 2021 11:27
Operator : JC/MD
Sample : VX1101MBL06
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Nov 02 03:30:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102821WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 02 03:29:47 2021
Response via : Initial Calibration

