

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110122\
 Data File : VX032509.D
 Acq On : 01 Nov 2022 13:38
 Operator : JC/MD
 Sample : N5391-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31846DL

Quant Time: Nov 02 05:29:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	216334	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188110	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87549	48.079	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.160%
35) Dibromofluoromethane	5.391	113	70134	45.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.900%
50) Toluene-d8	8.647	98	256275	47.761	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.079	95	97668	46.597	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.200%
Target Compounds						
16) Acetone	2.386	43	130449	197.103	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110122\
Data File : VX032509.D
Acq On : 01 Nov 2022 13:38
Operator : JC/MD
Sample : N5391-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31846DL

Quant Time: Nov 02 05:29:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

