

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110122\
 Data File : VX032515.D
 Acq On : 01 Nov 2022 15:58
 Operator : JC/MD
 Sample : N5390-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 02 05:31:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83123	48.629	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.260%
35) Dibromofluoromethane	5.385	113	65503	45.183	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.360%
50) Toluene-d8	8.647	98	240762	47.761	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.079	95	91141	46.285	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110122\
Data File : VX032515.D
Acq On : 01 Nov 2022 15:58
Operator : JC/MD
Sample : N5390-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 02 05:31:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

