

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110122\
 Data File : VX032517.D
 Acq On : 01 Nov 2022 16:45
 Operator : JC/MD
 Sample : N4955-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2022

Quant Time: Nov 02 05:31:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175095	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73546	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73605	48.284	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.560%
35) Dibromofluoromethane	5.391	113	59875	47.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.880%
50) Toluene-d8	8.653	98	200531	46.174	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.340%
62) 4-Bromofluorobenzene	11.079	95	74945	44.177	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110122\
Data File : VX032517.D
Acq On : 01 Nov 2022 16:45
Operator : JC/MD
Sample : N4955-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT4-2022

Quant Time: Nov 02 05:31:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

