



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110122\
Data File : VX032500.D
Acq On : 01 Nov 2022 08:47
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Title : SW846 8260
Last Update : Thu Oct 27 08:28:06 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	7247	PASS
75	95	30	60	56.3	19720	PASS
95	95	100	100	100.0	35011	PASS
96	95	5	9	6.3	2218	PASS
173	174	0.00	2	1.1	298	PASS
174	95	50	100	79.9	27968	PASS
175	174	5	9	7.4	2057	PASS
176	174	95	101	95.2	26616	PASS
177	176	5	9	6.8	1817	PASS