

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110218\
 Data File : VX005756.D
 Acq On : 02 Nov 2018 15:08
 Operator : JC/MD
 Sample : PB114401ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114401ZHE#07

Quant Time: Nov 03 02:11:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	108975	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.87	114	185064	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	168624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	70391	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	88669	55.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.30%	
35) Dibromofluoromethane	5.51	113	65254	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
50) Toluene-d8	8.72	98	234966	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
62) 4-Bromofluorobenzene	11.14	95	71555	43.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.28%	
Target Compounds						
18) Methyl Acetate	2.78	43	6781	2.966	ug/l	95
20) Methylene Chloride	2.86	84	1377	1.002	ug/l #	82
43) Isopropyl Acetate	6.46	43	95998	26.293	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110218\
Data File : VX005756.D
Acq On : 02 Nov 2018 15:08
Operator : JC/MD
Sample : PB114401ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114401ZHE#07

Quant Time: Nov 03 02:11:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

