

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110218\
 Data File : VX005759.D
 Acq On : 02 Nov 2018 16:29
 Operator : JC/MD
 Sample : PB114401ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114401ZHE#10

Quant Time: Nov 03 02:16:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	108188	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.87	114	179068	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	165936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	69262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	83805	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
35) Dibromofluoromethane	5.51	113	63606	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
50) Toluene-d8	8.72	98	228689	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
62) 4-Bromofluorobenzene	11.14	95	70402	43.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.74%	
Target Compounds						
18) Methyl Acetate	2.78	43	6524	2.874	ug/l	100
43) Isopropyl Acetate	6.46	43	97304	27.544	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110218\
Data File : VX005759.D
Acq On : 02 Nov 2018 16:29
Operator : JC/MD
Sample : PB114401ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114401ZHE#10

Quant Time: Nov 03 02:16:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

