

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110218\
 Data File : VX005760.D
 Acq On : 02 Nov 2018 16:56
 Operator : JC/MD
 Sample : PB114401ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114401ZHE#11

Quant Time: Nov 03 02:23:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	110983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	184171	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	164489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	69557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	83055	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
35) Dibromofluoromethane	5.51	113	62981	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	8.72	98	225949	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	11.14	95	70469	42.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.38%	
Target Compounds						
18) Methyl Acetate	2.78	43	5912	2.539	ug/l	97
20) Methylene Chloride	2.86	84	1525	1.090	ug/l	96
43) Isopropyl Acetate	6.46	43	93330	25.687	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110218\
Data File : VX005760.D
Acq On : 02 Nov 2018 16:56
Operator : JC/MD
Sample : PB114401ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114401ZHE#11

Quant Time: Nov 03 02:23:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

