

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110218\
 Data File : VX005764.D
 Acq On : 02 Nov 2018 18:44
 Operator : JC/MD
 Sample : PB114401ZHE#15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114401ZHE#15

Quant Time: Nov 03 02:53:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	107576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	179784	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	165760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	69977	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	88419	56.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.44%	
35) Dibromofluoromethane	5.50	113	66894	53.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.48%	
50) Toluene-d8	8.71	98	227236	52.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.10%	
62) 4-Bromofluorobenzene	11.14	95	71027	44.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.16%	
Target Compounds						
18) Methyl Acetate	2.78	43	6951	3.080	ug/l	96
43) Isopropyl Acetate	6.47	43	93291	26.302	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110218\
Data File : VX005764.D
Acq On : 02 Nov 2018 18:44
Operator : JC/MD
Sample : PB114401ZHE#15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114401ZHE#15

Quant Time: Nov 03 02:53:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

