

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110218\
 Data File : VX005765.D
 Acq On : 02 Nov 2018 19:11
 Operator : JC/MD
 Sample : J5776-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NJ-WHITE

Quant Time: Nov 03 02:55:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	110097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	193501	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	166889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	70116	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	88642	55.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.14%	
35) Dibromofluoromethane	5.50	113	66090	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	8.72	98	230765	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
62) 4-Bromofluorobenzene	11.14	95	71077	40.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.96%	
Target Compounds						
18) Methyl Acetate	2.78	43	6425	2.781	ug/l	97
20) Methylene Chloride	2.85	84	2165	1.560	ug/l	93
43) Isopropyl Acetate	6.47	43	92070	24.118	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110218\
Data File : VX005765.D
Acq On : 02 Nov 2018 19:11
Operator : JC/MD
Sample : J5776-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NJ-WHITE

Quant Time: Nov 03 02:55:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

