

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110218\
 Data File : VX005767.D
 Acq On : 02 Nov 2018 20:04
 Operator : JC/MD
 Sample : J5767-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-110118-B

Quant Time: Nov 03 02:59:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	107322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	168134	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	154765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	64507	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	80142	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
35) Dibromofluoromethane	5.51	113	63024	54.14	ug/l	0.01
Spiked Amount 50.000			Recovery	=	108.28%	
50) Toluene-d8	8.72	98	214417	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
62) 4-Bromofluorobenzene	11.14	95	65669	43.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.16%	
Target Compounds						
18) Methyl Acetate	2.78	43	6820	3.029	ug/l	95
20) Methylene Chloride	2.85	84	9956	7.358	ug/l	93
43) Isopropyl Acetate	6.46	43	85338	25.727	ug/l	99
68) m/p-Xylenes	10.36	106	2444	1.172	ug/l	88
95) Naphthalene	13.83	128	22875	5.085	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110218\
Data File : VX005767.D
Acq On : 02 Nov 2018 20:04
Operator : JC/MD
Sample : J5767-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-110118-B

Quant Time: Nov 03 02:59:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

