

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110218\
 Data File : VX005769.D
 Acq On : 02 Nov 2018 20:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 03 03:03:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.578	0.605	-4.7	101	0.00
3 P	Chloromethane	0.681	0.663	2.6	104	0.00
4 C	Vinyl Chloride	0.709	0.678	4.4#	98	0.00
5 T	Bromomethane	0.880	0.639	27.4#	84	0.00
6 T	Chloroethane	0.515	0.419	18.6	103	0.00
7 T	Trichlorofluoromethane	0.910	0.858	5.7	101	0.00
8 T	Diethyl Ether	0.409	0.372	9.0	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.553	0.505	8.7	101	0.00
10 T	Methyl Iodide	0.493	0.538	-9.1	104	0.00
11 T	Tert butyl alcohol	0.199	0.198	0.5	106	0.00
12 CM	1,1-Dichloroethene	0.539	0.476	11.7#	100	0.00
13 T	Acrolein	0.130	0.085	34.6#	65	0.00
14 T	Allyl chloride	1.114	0.990	11.1	96	0.00
15 T	Acrylonitrile	0.437	0.433	0.9	106	0.00
16 T	Acetone	0.416	0.377	9.4	103	0.00
17 T	Carbon Disulfide	1.655	1.386	16.3	95	0.00
18 T	Methyl Acetate	1.049	1.050	-0.1	113	0.00
19 T	Methyl tert-butyl Ether	1.864	1.762	5.5	100	0.00
20 T	Methylene Chloride	0.630	0.586	7.0	104	0.00
21 T	trans-1,2-Dichloroethene	0.570	0.504	11.6	100	0.00
22 T	Diisopropyl ether	2.066	1.952	5.5	102	0.00
23 T	Vinyl Acetate	1.799	1.741	3.2	100	0.00
24 P	1,1-Dichloroethane	1.168	1.094	6.3	102	0.00
25 T	2-Butanone	0.617	0.598	3.1	103	0.01
26 T	2,2-Dichloropropane	0.873	0.648	25.8#	81	0.00
27 T	cis-1,2-Dichloroethene	0.640	0.586	8.4	101	0.01
28 T	Bromochloromethane	0.570	0.537	5.8	104	0.00
29 T	Tetrahydrofuran	0.376	0.382	-1.6	106	0.01
30 C	Chloroform	1.142	1.067	6.6#	103	0.00
31 T	Cyclohexane	0.977	0.891	8.8	95	-0.01
32 T	1,1,1-Trichloroethane	0.914	0.861	5.8	101	0.00
33 S	1,2-Dichloroethane-d4	0.731	0.694	5.1	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.346	0.348	-0.6	113	0.00
36 T	1,1-Dichloropropene	0.533	0.496	6.9	98	0.00
37 T	Ethyl Acetate	0.680	0.684	-0.6	107	0.00
38 T	Carbon Tetrachloride	0.481	0.466	3.1	101	0.00
39 T	Methylcyclohexane	0.556	0.509	8.5	93	0.00
40 TM	Benzene	1.585	1.512	4.6	99	0.00
41 T	Methacrylonitrile	0.367	0.359	2.2	100	0.00
42 TM	1,2-Dichloroethane	0.586	0.565	3.6	101	0.00
43 T	Isopropyl Acetate	0.986	0.973	1.3	100	0.00
44 TM	Trichloroethene	0.367	0.342	6.8	100	0.00
45 C	1,2-Dichloropropane	0.406	0.389	4.2#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX110218\
 Data File : VX005769.D
 Acq On : 02 Nov 2018 20:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 03 03:03:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.260	0.242	6.9	102	0.00
47 T	Bromodichloromethane	0.487	0.473	2.9	100	0.00
48 T	Methyl methacrylate	0.477	0.482	-1.0	101	0.00
49 T	1,4-Dioxane	0.011	0.011	0.0	107	0.00
50 S	Toluene-d8	1.214	1.189	2.1	106	0.00
51 T	4-Methyl-2-Pentanone	0.673	0.699	-3.9	104	0.00
52 CM	Toluene	0.857	0.839	2.1#	97	0.00
53 T	t-1,3-Dichloropropene	0.528	0.530	-0.4	98	0.00
54 T	cis-1,3-Dichloropropene	0.565	0.557	1.4	98	0.00
55 T	1,1,2-Trichloroethane	0.372	0.405	-8.9	110	0.00
56 T	Ethyl methacrylate	0.562	0.607	-8.0	103	0.00
57 T	1,3-Dichloropropane	0.639	0.710	-11.1	116	0.00
58 T	2-Chloroethyl Vinyl ether	0.318	0.344	-8.2	102	0.00
59 T	2-Hexanone	0.547	0.588	-7.5	106	0.00
60 T	Dibromochloromethane	0.360	0.387	-7.5	108	0.00
61 T	1,2-Dibromoethane	0.379	0.391	-3.2	99	0.00
62 S	4-Bromofluorobenzene	0.448	0.443	1.1	107	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.365	0.420	-15.1	116	0.00
65 PM	Chlorobenzene	1.062	1.024	3.6	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.373	-2.5	106	0.00
67 C	Ethyl Benzene	1.813	1.886	-4.0#	104	0.00
68 T	m/p-Xylenes	0.674	0.690	-2.4	101	0.00
69 T	o-Xylene	0.644	0.657	-2.0	100	0.00
70 T	Styrene	1.070	1.130	-5.6	102	0.00
71 P	Bromoform	0.295	0.296	-0.3	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
73 T	Isopropylbenzene	3.299	3.378	-2.4	101	0.00
74 T	N-amyl acetate	1.756	1.770	-0.8	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.369	1.324	3.3	102	0.00
76 T	1,2,3-Trichloropropane	1.245	1.227	1.4	104	0.00
77 T	Bromobenzene	0.818	0.805	1.6	102	0.00
78 T	n-propylbenzene	4.013	4.066	-1.3	100	0.00
79 T	2-Chlorotoluene	2.397	2.396	0.0	101	0.00
80 T	1,3,5-Trimethylbenzene	2.801	2.876	-2.7	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.385	0.361	6.2	95	0.00
82 T	4-Chlorotoluene	2.861	2.831	1.0	100	0.00
83 T	tert-Butylbenzene	2.672	2.698	-1.0	100	0.00
84 T	1,2,4-Trimethylbenzene	2.828	2.953	-4.4	101	0.00
85 T	sec-Butylbenzene	3.428	3.464	-1.1	101	0.00
86 T	p-Isopropyltoluene	2.916	2.970	-1.9	101	0.00
87 T	1,3-Dichlorobenzene	1.582	1.520	3.9	101	0.00
88 T	1,4-Dichlorobenzene	1.700	1.556	8.5	103	0.00
89 T	n-Butylbenzene	2.942	2.831	3.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110218\
Data File : VX005769.D
Acq On : 02 Nov 2018 20:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 03 03:03:31 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.517	0.498	3.7	99	0.00
91 T	1,2-Dichlorobenzene	1.653	1.564	5.4	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.321	0.326	-1.6	107	0.00
93 T	1,2,4-Trichlorobenzene	1.060	1.047	1.2	102	0.00
94 T	Hexachlorobutadiene	0.513	0.480	6.4	102	0.00
95 T	Naphthalene	3.487	3.814	-9.4	103	0.00
96 T	1,2,3-Trichlorobenzene	1.079	1.089	-0.9	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6