


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110218\
Data File : VX005743.D
Acq On : 02 Nov 2018 09:05
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Title : SW846 8260
Last Update : Wed Oct 31 18:23:06 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	3109	PASS
75	95	30	60	51.9	7644	PASS
95	95	100	100	100.0	14726	PASS
96	95	5	9	6.9	1015	PASS
173	174	0.00	2	1.4	169	PASS
174	95	50	100	84.6	12451	PASS
175	174	5	9	7.9	979	PASS
176	174	95	101	97.9	12191	PASS
177	176	5	9	6.4	786	PASS