

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110220\
 Data File : VX019206.D
 Acq On : 02 Nov 2020 11:39
 Operator : JC/MD
 Sample : VSTD05034
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 02 12:55:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXLM110220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 02 12:48:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	458909	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	427483	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.06	152	209159	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	193149	63.35	ug/L	0.00
7) Chloroethane-d5	1.70	69	148324	64.39	ug/L	0.00
11) 1,1-Dichloroethene-d2	2.35	63	326170	52.61	ug/L	0.00
21) 2-Butanone-d5	4.53	46	207195	102.40	ug/L	0.00
24) Chloroform-d	5.14	84	317434	54.38	ug/L	0.00
26) 1,2-Dichloroethane-d4	6.03	65	210547	47.92	ug/L	0.00
32) Benzene-d6	6.05	84	650106	58.64	ug/L	0.00
36) 1,2-Dichloropropane-d6	7.37	67	191451	55.46	ug/L	0.00
41) Toluene-d8	8.70	98	627957	57.91	ug/L	0.00
43) trans-1,3-Dichloropropene-	8.99	79	102651	51.77	ug/L	0.00
47) 2-Hexanone-d5	9.43	63	168187	117.23	ug/L	0.00
56) 1,1,2,2-Tetrachloroethane-	11.23	84	236374	51.14	ug/L	0.00
66) 1,2-Dichlorobenzene-d4	12.36	152	206429	51.19	ug/L	0.00

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.18	85	149639	37.867	ug/L	99
3) Chloromethane	1.31	50	153134	50.760	ug/L	100
5) Vinyl chloride	1.40	62	171306	51.057	ug/L	99
6) Bromomethane	1.64	94	113906	63.701	ug/L	95
8) Chloroethane	1.72	64	110892	56.257	ug/L	98
9) Trichlorofluoromethane	1.92	101	257306	45.405	ug/L	100
10) 1,1,2-Trichloro-1,2,2-trif	2.37	101	147188	47.775	ug/L	95
12) 1,1-Dichloroethene	2.36	96	141957	49.108	ug/L	82
13) Acetone	2.42	43	132376	88.942	ug/L	94
14) Carbon disulfide	2.56	76	427780	48.666	ug/L	99
15) Methyl Acetate	2.75	43	140891	44.610	ug/L #	88
16) Methylene chloride	2.84	84	158390	50.012	ug/L	92
17) trans-1,2-Dichloroethene	3.14	96	154056	50.645	ug/L	97
18) Methyl tert-butyl Ether	3.17	73	488596	49.861	ug/L	97
19) 1,1-Dichloroethane	3.67	63	273730	48.402	ug/L	100
20) cis-1,2-Dichloroethene	4.56	96	168804	50.747	ug/L	93
22) 2-Butanone	4.64	43	207125	91.655	ug/L	91
23) Bromochloromethane	4.98	128	84382	46.101	ug/L	92
25) Chloroform	5.17	83	288703	45.663	ug/L	99
27) 1,2-Dichloroethane	6.16	62	230303	45.514	ug/L	97
29) Cyclohexane	5.55	56	255103	54.390	ug/L	96
30) 1,1,1-Trichloroethane	5.46	97	264893	46.116	ug/L	99
31) Carbon tetrachloride	5.75	117	227216	43.790	ug/L	99
33) Benzene	6.11	78	632057	52.709	ug/L	100
34) Trichloroethene	7.18	95	170980	51.101	ug/L	97
35) Methylcyclohexane	7.44	83	272329	56.006	ug/L	96
37) 1,2-Dichloropropane	7.49	63	157315	49.710	ug/L	99
38) Bromodichloromethane	7.87	83	217678	47.905	ug/L	99
39) cis-1,3-Dichloropropene	8.42	75	262503	51.386	ug/L	100
40) 4-Methyl-2-pentanone	8.62	43	397986	94.816	ug/L #	92

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110220\
 Data File : VX019206.D
 Acq On : 02 Nov 2020 11:39
 Operator : JC/MD
 Sample : VSTD05034
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 02 12:55:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXLM110220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 02 12:48:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

42) Toluene	8.76	91	699808	54.322	ug/L	99
44) trans-1,3-Dichloropropene	9.02	75	258772	49.776	ug/L	96
45) 1,1,2-Trichloroethane	9.20	97	156144	50.936	ug/L	99
46) Tetrachloroethene	9.32	164	128489	46.662	ug/L	97
48) 2-Hexanone	9.47	43	313944	94.958	ug/L #	92
49) Dibromochloromethane	9.56	129	174346	45.136	ug/L	98
50) 1,2-Dibromoethane	9.65	107	164702	48.851	ug/L	96
51) Chlorobenzene	10.12	112	439123	50.751	ug/L	100
52) Ethylbenzene	10.23	91	767719	53.498	ug/L	99
53) m,p-Xylene	10.34	106	293794	53.751	ug/L	98
54) o-Xylene	10.68	106	283619	53.875	ug/L	99
55) Styrene	10.70	104	486289	53.653	ug/L	94
57) 1,1,2,2-Tetrachloroethane	11.25	83	222022	47.330	ug/L	99
59) Bromoform	10.84	173	121196	43.947	ug/L	96
60) Isopropylbenzene	11.00	105	763165	57.033	ug/L	99
61) 1,2,3-Trichloropropane	11.28	75	190682	50.536	ug/L	99
62) 1,3,5-Trimethylbenzene	11.49	105	631827	57.194	ug/L	100
63) 1,2,4-Trimethylbenzene	11.79	105	600934	53.405	ug/L	100
64) 1,3-Dichlorobenzene	12.01	146	322462	48.789	ug/L	96
65) 1,4-Dichlorobenzene	12.08	146	326208	48.870	ug/L	96
67) 1,2-Dichlorobenzene	12.37	146	312783	49.150	ug/L	94
68) 1,2-Dibromo-3-chloropropan	12.98	75	50643	45.758	ug/L	95
69) 1,3,5-Trichlorobenzene	13.15	180	221837	50.599	ug/L	97
70) 1,2,4-trichlorobenzene	13.63	180	208599	51.837	ug/L	95
71) Naphthalene	13.82	128	701575	55.857	ug/L	99
72) 1,2,3-Trichlorobenzene	14.00	180	205097	50.849	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110220\
Data File : VX019206.D
Acq On : 02 Nov 2020 11:39
Operator : JC/MD
Sample : VSTD05034
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Nov 02 12:55:03 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXLM110220WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 02 12:48:14 2020
Response via : Initial Calibration

