

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110222\
 Data File : VX032535.D
 Acq On : 02 Nov 2022 14:06
 Operator : JC/MD
 Sample : N5420-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-PT-IN02-11012022

Quant Time: Nov 03 03:11:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195061	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171810	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79402	48.217	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.440%
35) Dibromofluoromethane	5.385	113	63744	45.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.620%
50) Toluene-d8	8.653	98	232676	48.092	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.180%
62) 4-Bromofluorobenzene	11.079	95	87254	46.169	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110222\
Data File : VX032535.D
Acq On : 02 Nov 2022 14:06
Operator : JC/MD
Sample : N5420-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-PT-IN02-11012022

Quant Time: Nov 03 03:11:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

