

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX110222\
 Data File : VX032526.D
 Acq On : 02 Nov 2022 10:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 03 03:06:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	50.000	50.082	-0.2	80	0.00
3 P	Chloromethane	50.000	47.806	4.4	80	0.00
4 C	Vinyl Chloride	50.000	47.673	4.7#	78	0.00
5 T	Bromomethane	50.000	55.315	-10.6	92	0.00
6 T	Chloroethane	50.000	47.253	5.5	85	0.00
7 T	Trichlorofluoromethane	50.000	49.678	0.6	84	0.00
8 T	Diethyl Ether	50.000	50.303	-0.6	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.728	-1.5	89	0.00
10 T	Methyl Iodide	50.000	50.926	-1.9	87	0.00
11 T	Tert butyl alcohol	250.000	234.256	6.3	81	0.00
12 CM	1,1-Dichloroethene	50.000	48.991	2.0#	84	0.00
13 T	Acrolein	250.000	270.874	-8.3	97	0.00
14 T	Allyl chloride	50.000	50.117	-0.2	85	0.00
15 T	Acrylonitrile	250.000	235.704	5.7	84	0.00
16 T	Acetone	250.000	260.654	-4.3	97	0.00
17 T	Carbon Disulfide	50.000	48.425	3.2	81	0.00
18 T	Methyl Acetate	50.000	51.162	-2.3	88	0.00
19 T	Methyl tert-butyl Ether	50.000	51.335	-2.7	91	0.00
20 T	Methylene Chloride	50.000	48.511	3.0	89	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.733	4.5	84	0.00
22 T	Diisopropyl ether	50.000	49.973	0.1	88	0.00
23 T	Vinyl Acetate	250.000	255.185	-2.1	88	0.00
24 P	1,1-Dichloroethane	50.000	49.237	1.5	86	0.00
25 T	2-Butanone	250.000	240.996	3.6	86	0.00
26 T	2,2-Dichloropropane	50.000	53.716	-7.4	92	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.539	0.9	85	0.00
28 T	Bromochloromethane	50.000	47.746	4.5	87	0.00
29 T	Tetrahydrofuran	250.000	235.142	5.9	82	0.00
30 C	Chloroform	50.000	50.632	-1.3#	89	0.00
31 T	Cyclohexane	50.000	49.132	1.7	83	0.00
32 T	1,1,1-Trichloroethane	50.000	51.536	-3.1	88	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.012	-2.0	92	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	85	0.00
35 S	Dibromofluoromethane	50.000	50.973	-1.9	89	0.00
36 T	1,1-Dichloropropene	50.000	50.796	-1.6	87	0.00
37 T	Ethyl Acetate	50.000	47.485	5.0	87	0.00
38 T	Carbon Tetrachloride	50.000	52.374	-4.7	88	-0.01
39 T	Methylcyclohexane	50.000	51.614	-3.2	85	0.00
40 TM	Benzene	50.000	49.948	0.1	85	0.00
41 T	Methacrylonitrile	50.000	50.785	-1.6	87	-0.01
42 TM	1,2-Dichloroethane	50.000	53.390	-6.8	91	0.00
43 T	Isopropyl Acetate	50.000	51.341	-2.7	86	0.00
44 TM	Trichloroethene	50.000	50.246	-0.5	87	0.00
45 C	1,2-Dichloropropane	50.000	50.890	-1.8#	88	0.00
46 T	Dibromomethane	50.000	51.500	-3.0	89	0.00
47 T	Bromodichloromethane	50.000	53.733	-7.5	91	0.00
48 T	Methyl methacrylate	50.000	51.250	-2.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110222\
 Data File : VX032526.D
 Acq On : 02 Nov 2022 10:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 03 03:06:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	919.017	8.1	77	0.00
50 S	Toluene-d8	50.000	49.713	0.6	86	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.836	-0.7	85	0.00
52 CM	Toluene	50.000	50.508	-1.0#	85	0.00
53 T	t-1,3-Dichloropropene	50.000	56.228	-12.5	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.370	-8.7	90	0.00
55 T	1,1,2-Trichloroethane	50.000	51.548	-3.1	91	0.00
56 T	Ethyl methacrylate	50.000	53.921	-7.8	88	0.00
57 T	1,3-Dichloropropane	50.000	51.501	-3.0	89	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	261.414	-4.6	86	0.00
59 T	2-Hexanone	250.000	253.510	-1.4	85	0.00
60 T	Dibromochloromethane	50.000	54.000	-8.0	91	0.00
61 T	1,2-Dibromoethane	50.000	52.283	-4.6	89	0.00
62 S	4-Bromofluorobenzene	50.000	51.892	-3.8	88	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	85	0.00
64 T	Tetrachloroethene	50.000	50.953	-1.9	87	0.00
65 PM	Chlorobenzene	50.000	50.616	-1.2	88	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.041	-6.1	91	0.00
67 C	Ethyl Benzene	50.000	51.513	-3.0#	87	0.00
68 T	m/p-Xylenes	100.000	101.731	-1.7	86	0.00
69 T	o-Xylene	50.000	51.867	-3.7	87	0.00
70 T	Styrene	50.000	52.602	-5.2	88	0.00
71 P	Bromoform	50.000	55.278	-10.6	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	85	0.00
73 T	Isopropylbenzene	50.000	49.434	1.1	88	0.00
74 T	N-amyl acetate	50.000	50.882	-1.8	86	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.911	4.2	86	0.00
76 T	1,2,3-Trichloropropane	50.000	50.275	-0.5	87	0.00
77 T	Bromobenzene	50.000	49.356	1.3	89	0.00
78 T	n-propylbenzene	50.000	49.976	0.0	87	0.00
79 T	2-Chlorotoluene	50.000	49.168	1.7	87	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.250	-0.5	86	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.754	-5.5	89	0.00
82 T	4-Chlorotoluene	50.000	50.251	-0.5	88	0.00
83 T	tert-Butylbenzene	50.000	50.270	-0.5	88	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.799	-1.6	87	0.00
85 T	sec-Butylbenzene	50.000	50.919	-1.8	87	0.00
86 T	p-Isopropyltoluene	50.000	51.979	-4.0	88	0.00
87 T	1,3-Dichlorobenzene	50.000	49.975	0.0	88	0.00
88 T	1,4-Dichlorobenzene	50.000	48.953	2.1	88	0.00
89 T	n-Butylbenzene	50.000	52.829	-5.7	88	0.00
90 T	Hexachloroethane	50.000	53.914	-7.8	92	0.00
91 T	1,2-Dichlorobenzene	50.000	49.638	0.7	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.377	-0.8	86	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.511	-1.0	91	0.00
94 T	Hexachlorobutadiene	50.000	50.997	-2.0	93	0.00
95 T	Naphthalene	50.000	49.006	2.0	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110222\
Data File : VX032526.D
Acq On : 02 Nov 2022 10:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 03 03:06:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.297	1.4	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6