



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110222\
Data File : VX032525.D
Acq On : 02 Nov 2022 09:27
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Title : SW846 8260
Last Update : Thu Oct 27 08:28:06 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	4655	PASS
75	95	30	60	56.0	13096	PASS
95	95	100	100	100.0	23395	PASS
96	95	5	9	7.3	1704	PASS
173	174	0.00	2	0.4	73	PASS
174	95	50	100	79.6	18629	PASS
175	174	5	9	7.8	1445	PASS
176	174	95	101	95.8	17841	PASS
177	176	5	9	6.0	1078	PASS