

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110322\
 Data File : VX032552.D
 Acq On : 03 Nov 2022 12:47
 Operator : JC/MD
 Sample : N5437-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 596C-3

Quant Time: Nov 04 06:18:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116492	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	199092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91219	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85868	53.050	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.100%	
35) Dibromofluoromethane	5.385	113	69008	48.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.180%	
50) Toluene-d8	8.647	98	240723	48.748	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.500%	
62) 4-Bromofluorobenzene	11.079	95	96350	49.949	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.900%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110322\
Data File : VX032552.D
Acq On : 03 Nov 2022 12:47
Operator : JC/MD
Sample : N5437-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
596C-3

Quant Time: Nov 04 06:18:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

