

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110322\
 Data File : VX032547.D
 Acq On : 03 Nov 2022 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 04 06:16:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.577	0.597	-3.5	79	0.00
3 P	Chloromethane	0.488	0.490	-0.4	80	0.00
4 C	Vinyl Chloride	0.537	0.556	-3.5#	80	0.00
5 T	Bromomethane	0.429	0.510	-18.9	94	0.00
6 T	Chloroethane	0.360	0.392	-8.9	93	0.00
7 T	Trichlorofluoromethane	1.017	1.052	-3.4	84	0.00
8 T	Diethyl Ether	0.330	0.332	-0.6	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.537	0.558	-3.9	86	0.00
10 T	Methyl Iodide	0.558	0.659	-18.1	89	0.00
11 T	Tert butyl alcohol	0.101	0.081	19.8	66	0.01
12 CM	1,1-Dichloroethene	0.501	0.509	-1.6#	83	0.00
13 T	Acrolein	0.101	0.103	-2.0	87	0.00
14 T	Allyl chloride	0.799	0.821	-2.8	83	0.00
15 T	Acrylonitrile	0.274	0.254	7.3	79	0.00
16 T	Acetone	0.253	0.294	-16.2	103	0.00
17 T	Carbon Disulfide	1.194	1.212	-1.5	80	0.00
18 T	Methyl Acetate	0.821	0.815	0.7	82	0.00
19 T	Methyl tert-butyl Ether	1.762	1.770	-0.5	85	0.00
20 T	Methylene Chloride	0.579	0.574	0.9	87	0.00
21 T	trans-1,2-Dichloroethene	0.546	0.547	-0.2	84	0.00
22 T	Diisopropyl ether	1.729	1.778	-2.8	87	0.00
23 T	Vinyl Acetate	1.397	1.405	-0.6	82	0.00
24 P	1,1-Dichloroethane	0.994	1.010	-1.6	84	0.00
25 T	2-Butanone	0.391	0.383	2.0	83	0.00
26 T	2,2-Dichloropropane	0.791	0.868	-9.7	89	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.645	-2.2	83	0.00
28 T	Bromochloromethane	0.354	0.348	1.7	86	0.00
29 T	Tetrahydrofuran	0.249	0.223	10.4	74	0.00
30 C	Chloroform	1.091	1.147	-5.1#	88	0.00
31 T	Cyclohexane	0.842	0.858	-1.9	82	0.00
32 T	1,1,1-Trichloroethane	0.979	1.023	-4.5	85	0.00
33 S	1,2-Dichloroethane-d4	0.695	0.700	-0.7	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.357	0.366	-2.5	84	0.00
36 T	1,1-Dichloropropene	0.497	0.530	-6.6	85	0.00
37 T	Ethyl Acetate	0.517	0.465	10.1	77	0.00
38 T	Carbon Tetrachloride	0.551	0.612	-11.1	87	-0.01
39 T	Methylcyclohexane	0.588	0.635	-8.0	83	0.00
40 TM	Benzene	1.405	1.474	-4.9	84	0.00
41 T	Methacrylonitrile	0.263	0.271	-3.0	82	0.00
42 TM	1,2-Dichloroethane	0.575	0.620	-7.8	86	0.00
43 T	Isopropyl Acetate	0.794	0.810	-2.0	80	0.00
44 TM	Trichloroethene	0.396	0.412	-4.0	84	0.00
45 C	1,2-Dichloropropane	0.361	0.386	-6.9#	86	0.00
46 T	Dibromomethane	0.277	0.290	-4.7	85	0.00
47 T	Bromodichloromethane	0.524	0.583	-11.3	88	0.00
48 T	Methyl methacrylate	0.410	0.418	-2.0	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110322\
 Data File : VX032547.D
 Acq On : 03 Nov 2022 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 04 06:16:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	70	0.01
50 S	Toluene-d8	1.240	1.271	-2.5	83	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.491	1.8	78	0.00
52 CM	Toluene	0.917	0.982	-7.1#	85	0.00
53 T	t-1,3-Dichloropropene	0.535	0.608	-13.6	87	0.00
54 T	cis-1,3-Dichloropropene	0.572	0.645	-12.8	88	0.00
55 T	1,1,2-Trichloroethane	0.386	0.403	-4.4	86	0.00
56 T	Ethyl methacrylate	0.556	0.593	-6.7	82	0.00
57 T	1,3-Dichloropropane	0.624	0.657	-5.3	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.286	-6.7	82	0.00
59 T	2-Hexanone	0.371	0.382	-3.0	80	0.00
60 T	Dibromochloromethane	0.422	0.469	-11.1	88	0.00
61 T	1,2-Dibromoethane	0.409	0.434	-6.1	85	0.00
62 S	4-Bromofluorobenzene	0.484	0.518	-7.0	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.366	0.393	-7.4	86	0.00
65 PM	Chlorobenzene	1.120	1.183	-5.6	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.422	0.470	-11.4	89	0.00
67 C	Ethyl Benzene	2.003	2.156	-7.6#	85	0.00
68 T	m/p-Xylenes	0.778	0.845	-8.6	85	0.00
69 T	o-Xylene	0.756	0.812	-7.4	84	0.00
70 T	Styrene	1.257	1.389	-10.5	86	0.00
71 P	Bromoform	0.333	0.377	-13.2	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.962	4.077	-2.9	85	0.00
74 T	N-amyl acetate	1.497	1.515	-1.2	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.272	1.255	1.3	83	0.00
76 T	1,2,3-Trichloropropane	1.112	1.137	-2.2	83	0.00
77 T	Bromobenzene	0.971	0.989	-1.9	86	0.00
78 T	n-propylbenzene	4.565	4.809	-5.3	86	0.00
79 T	2-Chlorotoluene	2.734	2.843	-4.0	86	0.00
80 T	1,3,5-Trimethylbenzene	3.386	3.597	-6.2	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.350	0.359	-2.6	81	0.00
82 T	4-Chlorotoluene	3.199	3.385	-5.8	87	0.00
83 T	tert-Butylbenzene	3.381	3.580	-5.9	86	0.00
84 T	1,2,4-Trimethylbenzene	3.368	3.595	-6.7	86	0.00
85 T	sec-Butylbenzene	4.068	4.383	-7.7	86	0.00
86 T	p-Isopropyltoluene	3.452	3.774	-9.3	86	0.00
87 T	1,3-Dichlorobenzene	1.832	1.940	-5.9	88	0.00
88 T	1,4-Dichlorobenzene	1.901	1.949	-2.5	86	0.00
89 T	n-Butylbenzene	3.003	3.387	-12.8	88	0.00
90 T	Hexachloroethane	0.564	0.630	-11.7	89	0.00
91 T	1,2-Dichlorobenzene	1.807	1.888	-4.5	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.262	6.1	75	0.00
93 T	1,2,4-Trichlorobenzene	1.178	1.227	-4.2	88	0.00
94 T	Hexachlorobutadiene	0.492	0.525	-6.7	91	0.00
95 T	Naphthalene	3.925	3.823	2.6	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110322\
Data File : VX032547.D
Acq On : 03 Nov 2022 10:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 04 06:16:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.197	1.213	-1.3	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6