

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX110322\
 Data File : VX032547.D
 Acq On : 03 Nov 2022 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 04 06:16:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	50.000	51.771	-3.5	79	0.00
3 P	Chloromethane	50.000	50.264	-0.5	80	0.00
4 C	Vinyl Chloride	50.000	51.828	-3.7#	80	0.00
5 T	Bromomethane	50.000	59.462	-18.9	94	0.00
6 T	Chloroethane	50.000	54.375	-8.8	93	0.00
7 T	Trichlorofluoromethane	50.000	51.715	-3.4	84	0.00
8 T	Diethyl Ether	50.000	50.406	-0.8	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.960	-3.9	86	0.00
10 T	Methyl Iodide	50.000	54.961	-9.9	89	0.00
11 T	Tert butyl alcohol	250.000	200.902	19.6	66	0.01
12 CM	1,1-Dichloroethene	50.000	50.713	-1.4#	83	0.00
13 T	Acrolein	250.000	255.886	-2.4	87	0.00
14 T	Allyl chloride	50.000	51.384	-2.8	83	0.00
15 T	Acrylonitrile	250.000	231.740	7.3	79	0.00
16 T	Acetone	250.000	290.745	-16.3	103	0.00
17 T	Carbon Disulfide	50.000	50.755	-1.5	80	0.00
18 T	Methyl Acetate	50.000	49.600	0.8	82	0.00
19 T	Methyl tert-butyl Ether	50.000	50.230	-0.5	85	0.00
20 T	Methylene Chloride	50.000	49.596	0.8	87	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.120	-0.2	84	0.00
22 T	Diisopropyl ether	50.000	51.411	-2.8	87	0.00
23 T	Vinyl Acetate	250.000	251.437	-0.6	82	0.00
24 P	1,1-Dichloroethane	50.000	50.789	-1.6	84	0.00
25 T	2-Butanone	250.000	244.798	2.1	83	0.00
26 T	2,2-Dichloropropane	50.000	54.929	-9.9	89	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.145	-2.3	83	0.00
28 T	Bromochloromethane	50.000	49.117	1.8	86	0.00
29 T	Tetrahydrofuran	250.000	223.186	10.7	74	0.00
30 C	Chloroform	50.000	52.571	-5.1#	88	0.00
31 T	Cyclohexane	50.000	50.932	-1.9	82	0.00
32 T	1,1,1-Trichloroethane	50.000	52.214	-4.4	85	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.351	-0.7	86	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	80	0.00
35 S	Dibromofluoromethane	50.000	51.260	-2.5	84	0.00
36 T	1,1-Dichloropropene	50.000	53.343	-6.7	85	0.00
37 T	Ethyl Acetate	50.000	45.012	10.0	77	0.00
38 T	Carbon Tetrachloride	50.000	55.577	-11.2	87	-0.01
39 T	Methylcyclohexane	50.000	54.038	-8.1	83	0.00
40 TM	Benzene	50.000	52.443	-4.9	84	0.00
41 T	Methacrylonitrile	50.000	51.462	-2.9	82	0.00
42 TM	1,2-Dichloroethane	50.000	54.001	-8.0	86	0.00
43 T	Isopropyl Acetate	50.000	51.035	-2.1	80	0.00
44 TM	Trichloroethene	50.000	52.128	-4.3	84	0.00
45 C	1,2-Dichloropropane	50.000	53.480	-7.0#	86	0.00
46 T	Dibromomethane	50.000	52.477	-5.0	85	0.00
47 T	Bromodichloromethane	50.000	55.609	-11.2	88	0.00
48 T	Methyl methacrylate	50.000	50.955	-1.9	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX110322\
 Data File : VX032547.D
 Acq On : 03 Nov 2022 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 04 06:16:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	890.020	11.0	70	0.01
50 S	Toluene-d8	50.000	51.230	-2.5	83	0.00
51 T	4-Methyl-2-Pentanone	250.000	245.400	1.8	78	0.00
52 CM	Toluene	50.000	53.545	-7.1#	85	0.00
53 T	t-1,3-Dichloropropene	50.000	56.793	-13.6	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	56.377	-12.8	88	0.00
55 T	1,1,2-Trichloroethane	50.000	52.127	-4.3	86	0.00
56 T	Ethyl methacrylate	50.000	53.380	-6.8	82	0.00
57 T	1,3-Dichloropropane	50.000	52.685	-5.4	85	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	266.900	-6.8	82	0.00
59 T	2-Hexanone	250.000	257.294	-2.9	80	0.00
60 T	Dibromochloromethane	50.000	55.617	-11.2	88	0.00
61 T	1,2-Dibromoethane	50.000	53.071	-6.1	85	0.00
62 S	4-Bromofluorobenzene	50.000	53.513	-7.0	85	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	79	0.00
64 T	Tetrachloroethene	50.000	53.680	-7.4	86	0.00
65 PM	Chlorobenzene	50.000	52.802	-5.6	86	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.734	-11.5	89	0.00
67 C	Ethyl Benzene	50.000	53.815	-7.6#	85	0.00
68 T	m/p-Xylenes	100.000	108.599	-8.6	85	0.00
69 T	o-Xylene	50.000	53.698	-7.4	84	0.00
70 T	Styrene	50.000	55.280	-10.6	86	0.00
71 P	Bromoform	50.000	56.590	-13.2	87	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	79	0.00
73 T	Isopropylbenzene	50.000	51.450	-2.9	85	0.00
74 T	N-ethyl acetate	50.000	50.604	-1.2	80	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.335	1.3	83	0.00
76 T	1,2,3-Trichloropropane	50.000	51.120	-2.2	83	0.00
77 T	Bromobenzene	50.000	50.918	-1.8	86	0.00
78 T	n-propylbenzene	50.000	52.678	-5.4	86	0.00
79 T	2-Chlorotoluene	50.000	52.005	-4.0	86	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.108	-6.2	86	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.412	-2.8	81	0.00
82 T	4-Chlorotoluene	50.000	52.903	-5.8	87	0.00
83 T	tert-Butylbenzene	50.000	52.951	-5.9	86	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.365	-6.7	86	0.00
85 T	sec-Butylbenzene	50.000	53.870	-7.7	86	0.00
86 T	p-Isopropyltoluene	50.000	54.663	-9.3	86	0.00
87 T	1,3-Dichlorobenzene	50.000	52.937	-5.9	88	0.00
88 T	1,4-Dichlorobenzene	50.000	51.267	-2.5	86	0.00
89 T	n-Butylbenzene	50.000	56.391	-12.8	88	0.00
90 T	Hexachloroethane	50.000	55.861	-11.7	89	0.00
91 T	1,2-Dichlorobenzene	50.000	52.248	-4.5	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.917	6.2	75	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.063	-4.1	88	0.00
94 T	Hexachlorobutadiene	50.000	53.379	-6.8	91	0.00
95 T	Naphthalene	50.000	48.695	2.6	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110322\
Data File : VX032547.D
Acq On : 03 Nov 2022 10:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 04 06:16:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.663	-1.3	85	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6