

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048447.D
 Acq On : 03 Nov 2025 11:25
 Operator : JC/MD
 Sample : VX1103WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1103WBL01

Quant Time: Nov 04 00:08:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	160467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	268071	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	290044	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	117306	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	121597	52.690	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.380%
35) Dibromofluoromethane	5.367	113	91442	59.830	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	119.660%
50) Toluene-d8	8.629	98	306638	45.451	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.900%#
62) 4-Bromofluorobenzene	11.061	95	122008	48.343	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048447.D
Acq On : 03 Nov 2025 11:25
Operator : JC/MD
Sample : VX1103WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1103WBL01

Quant Time: Nov 04 00:08:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

