

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048452.D
 Acq On : 03 Nov 2025 13:37
 Operator : JC/MD
 Sample : Q3506-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 04 00:12:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	133484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	237584	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	252349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	106003	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	113845	59.303	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	118.600%#
35) Dibromofluoromethane	5.367	113	73074	53.947	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.900%
50) Toluene-d8	8.628	98	315943	52.839	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.680%
62) 4-Bromofluorobenzene	11.061	95	123303	55.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048452.D
Acq On : 03 Nov 2025 13:37
Operator : JC/MD
Sample : Q3506-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 04 00:12:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

