

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048453.D
 Acq On : 03 Nov 2025 13:58
 Operator : JC/MD
 Sample : Q3506-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4RE

Quant Time: Nov 04 00:13:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	180470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	313503	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	234243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	116193	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	114016	43.929	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.860%
35) Dibromofluoromethane	5.373	113	86753	48.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.080%
50) Toluene-d8	8.634	98	298768	37.867	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	75.740%#
62) 4-Bromofluorobenzene	11.061	95	114814	38.900	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.800%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048453.D
Acq On : 03 Nov 2025 13:58
Operator : JC/MD
Sample : Q3506-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4RE

Quant Time: Nov 04 00:13:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

