

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048455.D
 Acq On : 03 Nov 2025 14:39
 Operator : JC/MD
 Sample : Q3511-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL-001

Quant Time: Nov 04 00:13:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	151059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	227617	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	210441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	122955	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	108164	49.789	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.580%
35) Dibromofluoromethane	5.373	113	71016	54.723	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.440%
50) Toluene-d8	8.634	98	267143	46.634	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.260%
62) 4-Bromofluorobenzene	11.061	95	106355	49.631	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.260%
Target Compounds						
16) Acetone	2.379	43	5697	11.570	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048455.D
Acq On : 03 Nov 2025 14:39
Operator : JC/MD
Sample : Q3511-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-001

Quant Time: Nov 04 00:13:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

