

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048456.D
 Acq On : 03 Nov 2025 14:59
 Operator : JC/MD
 Sample : Q3495-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-1025-1RE

Quant Time: Nov 04 00:14:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	146009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	263069	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	207013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	101418	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	100460	47.842	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.680%
35) Dibromofluoromethane	5.373	113	76564	51.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.100%
50) Toluene-d8	8.634	98	256435	38.732	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	77.460%#
62) 4-Bromofluorobenzene	11.067	95	97496	39.365	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	78.740%#
Target Compounds						
16) Acetone	2.386	43	1438	3.022	ug/l #	78
19) Methyl tert-butyl Ether	3.105	73	10356	1.903	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048456.D
Acq On : 03 Nov 2025 14:59
Operator : JC/MD
Sample : Q3495-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-1025-1RE

Quant Time: Nov 04 00:14:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

