

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048460.D
 Acq On : 03 Nov 2025 16:22
 Operator : JC/MD
 Sample : Q3494-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Field

Quant Time: Nov 04 00:15:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	147384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	222925	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	210225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	103668	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	108102	51.001	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	102.000%	
35) Dibromofluoromethane	5.373	113	75144	59.123	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	118.240%	
50) Toluene-d8	8.635	98	264437	47.133	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	94.260%	
62) 4-Bromofluorobenzene	11.061	95	107228	51.091	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	102.180%	
Target Compounds						
16) Acetone	2.373	43	3478	7.240	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048460.D
Acq On : 03 Nov 2025 16:22
Operator : JC/MD
Sample : Q3494-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Field

Quant Time: Nov 04 00:15:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

