

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048466.D
 Acq On : 03 Nov 2025 18:25
 Operator : JC/MD
 Sample : Q3521-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TANK-2025

Quant Time: Nov 04 00:18:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	149230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	266997	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	233886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	110763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	92816	43.248	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	86.500%		
35) Dibromofluoromethane	5.373	113	68847	45.227	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.460%		
50) Toluene-d8	8.635	98	299587	44.584	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	89.160%#		
62) 4-Bromofluorobenzene	11.061	95	117528	46.756	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.520%		
Target Compounds						
16) Acetone	2.380	43	2613	5.372	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048466.D
Acq On : 03 Nov 2025 18:25
Operator : JC/MD
Sample : Q3521-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TANK-2025

Quant Time: Nov 04 00:18:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

