

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
 Data File : VX048468.D
 Acq On : 03 Nov 2025 19:06
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Nov 04 00:19:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	152620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	275559	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	262307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	109484	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	115104	52.441	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.880%
35) Dibromofluoromethane	5.373	113	75342	47.956	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.920%
50) Toluene-d8	8.635	98	328666	47.392	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.780%
62) 4-Bromofluorobenzene	11.067	95	110063	42.425	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110325\
Data File : VX048468.D
Acq On : 03 Nov 2025 19:06
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Nov 04 00:19:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

