

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110419\
 Data File : VX013323.D
 Acq On : 04 Nov 2019 14:30
 Operator : JC/SP
 Sample : K5683-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OILY-WATERDL

Quant Time: Nov 05 01:01:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	757865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1158900	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1052861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	519043	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	428863	47.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.54%	
35) Dibromofluoromethane	5.48	113	351022	48.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.58%	
50) Toluene-d8	8.71	98	1400498	51.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.22%	
62) 4-Bromofluorobenzene	11.13	95	491760	49.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.44%	

Target Compounds						Qvalue
11) Tert butyl alcohol	3.01	59	19377	8.188	ug/l	100
16) Acetone	2.42	43	325384	72.657	ug/l	100
25) 2-Butanone	4.65	43	106066	16.145	ug/l	94
51) 4-Methyl-2-Pentanone	8.63	43	205957	16.410	ug/l	99
59) 2-Hexanone	9.48	43	30709	3.166	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110419\
Data File : VX013323.D
Acq On : 04 Nov 2019 14:30
Operator : JC/SP
Sample : K5683-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OILY-WATERDL

Quant Time: Nov 05 01:01:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

