

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110420\
 Data File : VX019239.D
 Acq On : 03 Nov 2020 17:48
 Operator : JC/MD
 Sample : PB132785TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132785TB

Quant Time: Nov 04 02:42:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	320200	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	561943	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	547153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	280795	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	212138	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
35) Dibromofluoromethane	5.46	113	171190	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
50) Toluene-d8	8.70	98	709470	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
62) 4-Bromofluorobenzene	11.12	95	276124	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
Target Compounds						
20) Methylene Chloride	2.84	84	3733	1.044	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110420\
Data File : VX019239.D
Acq On : 03 Nov 2020 17:48
Operator : JC/MD
Sample : PB132785TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
PB132785TB

Quant Time: Nov 04 02:42:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

