

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110420\
 Data File : VX019256.D
 Acq On : 04 Nov 2020 00:08
 Operator : JC/MD
 Sample : L4600-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT-17

Quant Time: Nov 04 03:08:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	316157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	560947	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	554892	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	286994	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	218393	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
35) Dibromofluoromethane	5.46	113	176015	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
50) Toluene-d8	8.70	98	714546	53.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.60%	
62) 4-Bromofluorobenzene	11.12	95	282200	56.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.20%	
Target Compounds						
16) Acetone	2.42	43	13882	9.869	ug/l	96
20) Methylene Chloride	2.84	84	11970	3.390	ug/l	89
43) Isopropyl Acetate	6.41	43	68931	8.560	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110420\
Data File : VX019256.D
Acq On : 04 Nov 2020 00:08
Operator : JC/MD
Sample : L4600-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CRT-17

Quant Time: Nov 04 03:08:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

