

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110420\
 Data File : VX019260.D
 Acq On : 04 Nov 2020 01:37
 Operator : JC/MD
 Sample : L4605-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT-01

Quant Time: Nov 04 03:15:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	316546	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	560589	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	556877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	283535	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	219073	53.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.32%	
35) Dibromofluoromethane	5.46	113	175920	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
50) Toluene-d8	8.70	98	708074	53.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.68%	
62) 4-Bromofluorobenzene	11.12	95	278578	55.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.84%	
Target Compounds						
16) Acetone	2.42	43	12486	8.866	ug/l	98
20) Methylene Chloride	2.84	84	12776	3.614	ug/l	97
43) Isopropyl Acetate	6.41	43	71805	8.922	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110420\
Data File : VX019260.D
Acq On : 04 Nov 2020 01:37
Operator : JC/MD
Sample : L4605-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CRT-01

Quant Time: Nov 04 03:15:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

