



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110421\
 Data File : VX025055.D
 Acq On : 03 Nov 2021 17:21
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X110421W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Wed Nov 03 23:52:56 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	6583	PASS
75	95	30	60	53.2	20156	PASS
95	95	100	100	100.0	37896	PASS
96	95	5	9	6.2	2334	PASS
173	174	0.00	2	1.6	390	PASS
174	95	50	100	65.6	24867	PASS
175	174	5	9	7.7	1907	PASS
176	174	95	101	96.5	23987	PASS
177	176	5	9	6.6	1575	PASS