

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
 Data File : VX048472.D
 Acq On : 04 Nov 2025 10:14
 Operator : JC/MD
 Sample : VX1104MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1104MBL01

Quant Time: Nov 05 01:42:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	110283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	243050	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	249398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	127802	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	80240	50.591	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.180%
35) Dibromofluoromethane	5.367	113	62180	44.872	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.740%
50) Toluene-d8	8.629	98	297225	48.591	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.061	95	129718	56.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
Data File : VX048472.D
Acq On : 04 Nov 2025 10:14
Operator : JC/MD
Sample : VX1104MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1104MBL01

Quant Time: Nov 05 01:42:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

