

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
 Data File : VX048473.D
 Acq On : 04 Nov 2025 10:36
 Operator : JC/MD
 Sample : VX1104WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1104WBL01

Quant Time: Nov 05 01:43:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	132762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	255401	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	268346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	146033	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	93163	48.794	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.580%
35) Dibromofluoromethane	5.367	113	73003	50.135	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.260%
50) Toluene-d8	8.634	98	309296	48.119	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.240%
62) 4-Bromofluorobenzene	11.061	95	140579	58.465	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
Data File : VX048473.D
Acq On : 04 Nov 2025 10:36
Operator : JC/MD
Sample : VX1104WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1104WBL01

Quant Time: Nov 05 01:43:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

