

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
 Data File : VX048479.D
 Acq On : 04 Nov 2025 13:21
 Operator : JC/MD
 Sample : Q3488-06 10X
 Misc : 3.51g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MOO-25-0311-0313

Quant Time: Nov 04 14:29:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	112086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	251158	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	258836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	125837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97353	60.394	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	120.780%#
35) Dibromofluoromethane	5.373	113	67407	47.074	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.140%
50) Toluene-d8	8.635	98	249277	39.437	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	78.880%#
62) 4-Bromofluorobenzene	11.061	95	110971	46.931	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.860%
Target Compounds						
18) Methyl Acetate	2.703	43	2265	1.854	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
Data File : VX048479.D
Acq On : 04 Nov 2025 13:21
Operator : JC/MD
Sample : Q3488-06 10X
Misc : 3.51g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MOO-25-0311-0313

Quant Time: Nov 04 14:29:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

