

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
 Data File : VX048484.D
 Acq On : 04 Nov 2025 15:11
 Operator : JC/MD
 Sample : Q3511-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL-002

Quant Time: Nov 05 03:27:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	142347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	286906	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	258868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	123426	60.291	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 120.580%#	
35) Dibromofluoromethane	5.373	113	83601	51.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.220%	
50) Toluene-d8	8.634	98	342920	47.492	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 94.980%	
62) 4-Bromofluorobenzene	11.061	95	136802	50.647	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.300%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	4123	8.886	ug/l	97
80) 1,3,5-Trimethylbenzene	11.433	105	1764	0.202	ug/l	94
84) 1,2,4-Trimethylbenzene	11.731	105	4269	0.495	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
Data File : VX048484.D
Acq On : 04 Nov 2025 15:11
Operator : JC/MD
Sample : Q3511-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-002

Quant Time: Nov 05 03:27:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

