

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
 Data File : VX048488.D
 Acq On : 04 Nov 2025 17:03
 Operator : JC/MD
 Sample : Q3537-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17-PURGE-WATER

Quant Time: Nov 05 03:29:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	128603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	264831	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	312284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	151015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	117367	63.458	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 126.920%#	
35) Dibromofluoromethane	5.373	113	83677	55.419	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.840%	
50) Toluene-d8	8.635	98	351651	52.760	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.520%	
62) 4-Bromofluorobenzene	11.061	95	153183	61.438	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 122.880%	
Target Compounds						Qvalue
16) Acetone	2.374	43	2683	6.401	ug/l	91
80) 1,3,5-Trimethylbenzene	11.433	105	2385	0.260	ug/l	91
84) 1,2,4-Trimethylbenzene	11.738	105	7048	0.777	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
Data File : VX048488.D
Acq On : 04 Nov 2025 17:03
Operator : JC/MD
Sample : Q3537-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17-PURGE-WATER

Quant Time: Nov 05 03:29:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

