

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110518\
 Data File : VX005773.D
 Acq On : 05 Nov 2018 12:57
 Operator : JC/MD
 Sample : VX1105WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1105WBL01

Quant Time: Nov 06 01:09:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	156506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	267222	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	232930	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	91610	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	115306	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
35) Dibromofluoromethane	5.50	113	89637	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	8.71	98	290531	44.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.54%	
62) 4-Bromofluorobenzene	11.14	95	94269	39.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.72%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110518\
Data File : VX005773.D
Acq On : 05 Nov 2018 12:57
Operator : JC/MD
Sample : VX1105WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1105WBL01

Quant Time: Nov 06 01:09:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

