

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110518\
 Data File : VX005789.D
 Acq On : 05 Nov 2018 20:04
 Operator : JC/MD
 Sample : J5816-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-4A-S

Quant Time: Nov 06 02:33:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	146222	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	238172	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	212746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	93252	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	109649	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
35) Dibromofluoromethane	5.50	113	83098	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
50) Toluene-d8	8.71	98	295596	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	11.14	95	94372	44.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.42%	
Target Compounds						
8) Diethyl Ether	2.19	74	4271	3.571	ug/l	97
16) Acetone	2.45	43	13211	2.804	ug/l	94
18) Methyl Acetate	2.78	43	5397	1.759	ug/l	96
20) Methylene Chloride	2.86	84	3429	1.860	ug/l	85
25) 2-Butanone	4.70	43	3536	1.961	ug/l	100
43) Isopropyl Acetate	6.46	43	108949	23.187	ug/l	98
95) Naphthalene	13.83	128	17073	2.626	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110518\
Data File : VX005789.D
Acq On : 05 Nov 2018 20:04
Operator : JC/MD
Sample : J5816-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-4A-S

Quant Time: Nov 06 02:33:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

