

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110518\
 Data File : VX005795.D
 Acq On : 05 Nov 2018 22:44
 Operator : JC/MD
 Sample : J5665-03ME
 Misc : 3.06g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DL-07AME

Quant Time: Nov 06 09:15:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	150018	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	244518	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	218070	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	106191	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	112487	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
35) Dibromofluoromethane	5.50	113	81582	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	304944	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
62) 4-Bromofluorobenzene	11.14	95	104872	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
Target Compounds						
18) Methyl Acetate	2.81	43	33586	10.670	ug/l #	91
39) Methylcyclohexane	7.45	83	29636	10.908	ug/l	96
67) Ethyl Benzene	10.25	91	16440	2.079	ug/l	97
68) m/p-Xylenes	10.36	106	38942	13.251	ug/l	97
69) o-Xylene	10.70	106	16358	5.820	ug/l	93
84) 1,2,4-Trimethylbenzene	11.81	105	12569	2.092	ug/l	98
95) Naphthalene	13.83	128	21638	2.922	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110518\
Data File : VX005795.D
Acq On : 05 Nov 2018 22:44
Operator : JC/MD
Sample : J5665-03ME
Misc : 3.06µg/10mL/100µL/5.0mL/MSVOA_X/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DL-07AME

Quant Time: Nov 06 09:15:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

