

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110524\
 Data File : VX043717.D
 Acq On : 05 Nov 2024 15:36
 Operator : JC/MD
 Sample : P4621-10RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BH6N6RE

Quant Time: Nov 06 02:04:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 06 01:38:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	169871	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	159466	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	69739	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	56245	62.152	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 124.300%			
7) Chloroethane-d5	1.648	69	27782	71.421	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 142.840%#			
11) 1,1-Dichloroethene-d2	2.294	65	23207	55.551	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 111.100%			
21) 2-Butanone-d5	4.465	46	88068	148.631	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 148.630%#			
24) Chloroform-d	5.056	84	116237	59.582	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 119.160%			
26) 1,2-Dichloroethane-d4	5.952	65	77802	62.777	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 125.560%#			
32) Benzene-d6	5.970	84	248868	60.363	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 120.720%			
36) 1,2-Dichloropropane-d6	7.305	67	77385	61.695	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 123.380%#			
41) Toluene-d8	8.647	98	218108	58.236	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 116.480%			
43) trans-1,3-Dichloroprop...	8.951	79	31950	53.263	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 106.520%			
47) 2-Hexanone-d5	9.384	63	64661	130.260	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 130.260%#			
56) 1,1,2,2-Tetrachloroeth...	11.189	84	101186	62.873	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 125.740%#			
66) 1,2-Dichlorobenzene-d4	12.323	152	80946	66.991	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 133.980%#			
Target Compounds					Qvalue	
13) Acetone	2.386	43	2151	3.903	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110524\
Data File : VX043717.D
Acq On : 05 Nov 2024 15:36
Operator : JC/MD
Sample : P4621-10RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BH6N6RE

Quant Time: Nov 06 02:04:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 06 01:38:34 2024
Response via : Initial Calibration

