

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110618\
 Data File : VX005812.D
 Acq On : 06 Nov 2018 16:38
 Operator : JC/MD
 Sample : PB114531ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114531ZHE#07

Quant Time: Nov 07 09:08:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	176881	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	281755	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	248194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	126219	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
35) Dibromofluoromethane	5.50	113	96771	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	8.71	98	340814	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
62) 4-Bromofluorobenzene	11.14	95	109775	43.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.94%	
Target Compounds						
18) Methyl Acetate	2.78	43	7002	1.887	ug/l #	91
20) Methylene Chloride	2.86	84	8592	3.853	ug/l	94
43) Isopropyl Acetate	6.46	43	144209	25.943	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110618\
Data File : VX005812.D
Acq On : 06 Nov 2018 16:38
Operator : JC/MD
Sample : PB114531ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114531ZHE#07

Quant Time: Nov 07 09:08:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

