

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110618\
 Data File : VX005813.D
 Acq On : 06 Nov 2018 17:05
 Operator : JC/MD
 Sample : PB114531ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114531ZHE#08

Quant Time: Nov 07 09:27:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	165839	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	298866	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	241163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	105170	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	124282	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
35) Dibromofluoromethane	5.50	113	94696	45.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.52%	
50) Toluene-d8	8.71	98	344809	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
62) 4-Bromofluorobenzene	11.14	95	106608	39.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.60%	
Target Compounds						
18) Methyl Acetate	2.78	43	8005	2.301	ug/l	92
20) Methylene Chloride	2.86	84	8636	4.130	ug/l	90
25) 2-Butanone	4.71	43	2182	1.067	ug/l #	87
43) Isopropyl Acetate	6.46	43	138773	23.536	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110618\
Data File : VX005813.D
Acq On : 06 Nov 2018 17:05
Operator : JC/MD
Sample : PB114531ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114531ZHE#08

Quant Time: Nov 07 09:27:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

